



Have questions? We're here to help!
[Click here to contact us](#)

[View Online Catalog](#)

Mask 3Ply - Kids Blue

SKU: HMS-3PLY-KIDS

Product Details

Price per Unit	<i>Request a quote for price</i>
Unit Size	Each
Sold in multiples of	3,000
Delivery	Delivers in 7-10 Days

Minimum Order Quantities (MOQ), delivery times, and availability are subject to change without notice. Please contact us directly for the most current pricing, availability and delivery time to your location.

How to Buy

Unless otherwise indicated, all products are on-the-ground in the United States. Ship date is dependent on individual Terms & Conditions. Delivery can be expected 3-10 days from ship date.

To request pricing and/or place an order, please click our form link below. To ensure accurate information on pricing and delivery, please use the corresponding product(s) SKU.

[Complete this form to Request Pricing or Place an Order](#)

About Our Products

The ASK team has family members on the front lines so our commitment is deep and personal. Working with a network of experienced and reliable distributors and FDA Approved and FDA Compliant manufacturers, we're working to support organizations of all sizes. **No BS.**

[PPE Resources](#)

Not another flash in the pan.

Commitment to transparency. ASK PPE is part of ASK Consulting Group LLC, a business advisory group founded more than 20 years ago to help companies grow. Early in the COVID-19 pandemic, while helping a client secure financing to acquire PPE, we experienced the type of wild-west, no-rules, free-for-all we haven't seen since marijuana was legalized. Questionable brokers, the sale of product that didn't exist and worst of all, deposits made on product that was never delivered. **In short, we saw lots of BS – business scams.**

[Learn More About ASK PPE](#)



ASK PPE

<https://askppe.com>

P: 602-524-9138 ext. 409

E: sales@askppe.com

Test Report

SL52025269904401TX

Date: June 30, 2020

Page 1 of 3

HUNAN JIALONG MEDICAL EQUIPMENT CO., LTD
BUILDING 9, AUDIO-ELECTRONIC DISPLAY AND COMMUNICATION INDUSTRIAL PARK, HIGH-TECH
INDUSTRIAL PARK, LENGSHUITAN DISTRICT, YONGZHOU CITY, HUNAN PROVINCE, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Disposable medical mask (ear mounted type)

SGS Internal Ref No. : SHHL2006521229MD
Style No. : Tape A (ear mounted)
Sample Color : (A)Blue
Lot No./Batch No. : Not provided
Other Information : Age Grading: Two Years

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Jun 08, 2020
Testing Period : Jun 08, 2020 - Jun 30, 2020
Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.
Testing Center

3rd Building No.889, Yishan Road, Xuhui District Shanghai, China 200233
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763
t (86-21) 61402666 f (86-21) 64958763

www.sgs.com
www.sgs.com.cn
e sgs.china@sgs.com

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial filtration efficiency (BFE)*

(EN 14683:2019 Annex B)

Sample: A

	1#	2#	3#	4#	5#
(BFE), %	99.8	99.9	99.9	99.9	99.8

Remark: Performance Requirement: Type I $\geq 95\%$, Type II $\geq 98\%$, Type IIR $\geq 98\%$

* This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation).

Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C)

Sample: A

Test number and location : 5 random areas for each specimen (face mask)
 Conditioning Parameters : Minimum of 4 hours at 21 \pm 5°C and 85 \pm 5% R.H.
 Test Area : 4.9 cm²
 Flow Rate : 8 l/min

	1#	2#	3#	4#	5#
Differential pressure ΔP (Pa/cm ²)	32	32	33	32	32

Remark: Performance Requirement: Type I <40 Pa/cm², Type II <40 Pa/cm², Type IIR <60 Pa/cm²



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
 Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

Clause 5.2.5 Microbial Cleanliness

(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Sample Number	Mask Weight(g)	Total Bioburden, (CFU/mask)	Total Bioburden, (CFU/g)
1#	3.47	<3	<0.86
2#	3.52	<3	<0.85
3#	3.47	<3	<0.86
4#	3.41	24	7.04
5#	3.46	3	0.87

Remark: Performance Requirement: Type I ≤ 30 CFU/g, Type II ≤ 30 CFU/g, Type IIR ≤ 30 CFU/g

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.
Testing Center

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
3rd Building No.889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com